

# *Look Busy*

e-Liquid Samples



*Analysis Report*  
(02-15-17)

**GC/FID Analysis (ANALYSIS CODE)**

Diacetyl

2,3-Pentanedione (aka Acetyl propionyl)



Dvine Laboratories, Inc  
Phone: 905-985-4520  
[www.dvineejuce.com](http://www.dvineejuce.com)  
175 Northport Road  
Port Perry, ON L9L1B2

I certify that to the best of my knowledge all analytical data presented in this report:

- Have been reviewed for completeness and accuracy
- Are legible and free of errors
- Have been conducted in conformity with approved agreement, and that all deviations and analytical problems are condensed in the appropriate descriptions

Prepared in PDF format by Dvine Laboratories



---

Signature

Review Performed by:

Kristen Cain

Report issued:

February 21st, 2017



# Summary of Results



# Narrative Summary

## Dvine Laboratories Narrative Summary

Client	Look Busy
Analysts	KC
Parameters	<b>GC/FID Analysis</b>

Client project	
Job #	02-15-17
# Samples	4

### Custody

The samples were prepared on site on February 14<sup>th</sup>, 2017 at Dvine Laboratories. They were then received at 9:00am (02-15-17) at 19.5°C by Kristen Cain of Dvine Laboratories. All samples had tamper proof caps still intact. Contents were kept in original sealed containers before and after sampling and were not accessed by anyone except the technologist, Kristen Cain.

### Analysis

The samples were analyzed for diacetyl, 2,3-pentanedione (aka acetyl propionyl) and nicotine, following the analytical procedures in Dvine Laboratories SOP

DVNL-ELT-001-V1.

A measured volume of sample was combined with a measured volume of Methanol to prepare the sample. Two Samples were prepared and 3 injections were completed. All data was then measured using HP ChemStation software. If a sample was confirmed to have either or both compounds after review, a new test was run. The new test was run with appropriate Blanks to ensure Column Condition (Zero left over amounts from other previous samples) and then compared with the previous report. Additionally, should both reports show differences then new samples were prepared and run based on 3 Blanks initially to ensure column Condition, and this provided final analysis to ensure accuracy of results.

The HP Model 5890 Series II, GC was equipped with a FID and an appropriate Restek column.



# Chain of Custody if Applicable



**This Is the Last Page of  
Report 02-15-17 by Dvine  
Laboratories**